



Ref. Certif. No.

**DK-75747-M1-UL**

**IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME**

**CB TEST CERTIFICATE**

Product

Battery chargers (Power Supply)

Name and address of the applicant

MASCOT ELECTRONICS AS  
MOSSEVEIEN 109  
GRESSVIK, N-1624 NORWAY

Name and address of the manufacturer

MASCOT ELECTRONICS AS  
MOSSEVEIEN 109  
GRESSVIK, N-1624 NORWAY

Name and address of the factory

MASCOT BALTIC OU  
TAEVAKIVI TN 15 TALLINN 13619 EE  
ESTONIA

Note: When more than one factory, please report on page 2

Additional Information on page 2

Ratings and principal characteristics

Input ratings: 100-240Vac, 50-60Hz, max. 0.25A.

Trademark / Brand (if any)



Customer's Testing Facility (CTF) Stage used

3745x, 3745Bx, 3845x, 3845Bx. See nomenclature breakdown in test report for details.

Model / Type Ref.

The report was revised to include technical modifications.

Additional information (if necessary may also be reported on page 2)

Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1-11:2015, IEC60601-1-6:2010/AMD1:2013,  
IEC 60601-1-6:2010, IEC 60601-1:2005/AMD1:2012,  
IEC 60601-1:2005

As shown in the Test Report Ref. No. which forms part of this Certificate

E356182-D1006-1/A1/C0-ULCB issued on 2021-03-22

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2021-04-06

Signature:

Original Issue Date: 2018-08-15

Jan-Erik Storgaard



Ref. Certif. No.

**DK-75747-M1-UL**

**Factories:**

Mascot Power Supplies (Ningbo) Co., Ltd  
No.128 Jinchuan Road, Zhenhai Ningbo 315221  
China

**Additional Information:**

The original report was modified to include following changes/additions:

- update of model names, nomenclatures, ratings, applicant and manufacturer addresses;
- removal of one of the factories;
- addition of two more resistors.

Additionally evaluated to EN 60601-1:2006/ A1:2013/ A12:2014, EN 60601-1-11:2015 and EN 60601-1-6:2010/ A1:2015; National Differences specified in the CB Test Report.

**Additional information (if necessary)**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

**Date:** 2021-04-06  
Original Issue Date: 2018-08-15

**Signature:** Jan-Erik Storgaard